
Function generator 0.1Hz-2MHz PM 5132

9445 051 32001

Instruction manual - Gerätehandbuch - Mode d'emploi et d'entretien 9499 453 00202 88 04 01/6



Function generator 0.1Hz-2MHz PM 5132

9445 051 32001

Instruction manual Gerätehandbuch Mode d'emploi et d'entretien 9499 453 00202 88 04 01/6





PHILIPS

Please note

In correspondence concerning this instrument, please quote the type number and serial number as given on the type plate.

Bitte beachten

Bei Schriftwechsel über dieses Gerät wird gebeten, die Typennummer und die Gerätenummer anzugeben. Diese befinden sich auf dem Typenschild an der Rückseite des Gerätes.

Noter s. v. p.

Dans votre correspondance et dans vos réclamations se rapportant à cet appareil, veuillez toujours indiquer le numéro de type et le numéro de série qui sont marqués sur la plaquette de caractéristiques.

Important

As the instrument is an electrical apparatus, it may be operated only by trained personnel. Maintenance and repairs may also be carried out only by qualified personnel.

Wichtig

Da das Gerät ein elektrisches Betriebsmittel ist, darf die Bedienung nur durch eingewiesenes Personal erfolgen. Wartung und Reparatur dürfen nur von geschultem, fach- und sachkundigem Personal durchgeführt werden.

Important

Comme l'instrument est un équipement électrique, le service doit être assuré par du personnel qualifié. De même, l'entretien et les réparations sont à confier aux personnes suffisamment qualifiées.



Philips GmbH - Hamburg - Germany - 1988

Ang

All rights are strictly reserved.

CONTENTS

١.	GENERAL	5
i.1.	Introduction	5
1.2.	Technical data	5
1.3.	Accessories	9
1.4.	Operating principle	10
	Operating principle	
2.	INSTALLATION INSTRUCTIONS	11
2.1.	Initial inspection	11
2.2.	Safety instructions	11
2.3.	Mains voltage setting and fuses	12
2.4.	Operating position of the instrument	12
2.5.	Dismantling the instrument	12
2.6.	Radio interference suppression	12
3.	OPERATING INSTRUCTIONS	13
3.1.	Controls and sockets	13
3.1. 3.2.	Operation	14
3.2.	Operation	
4.	SERVICE PART	41
4.1.	Circuit description	41
4.2.	Access to parts	45
4.3.	Check and adjustment	46
4.4.	Safety inspection and tests after repair and maintenance	49
4.5.	Spare parts	49
5.	FIGURES	
•		
	1. Block diagram	
	Front view Rear view	
	4. Unit 2, component lay-out	
	5. Front view, mechanical parts 6. Bottom view	
	7. Unit 1, component lay -out	
	8. Overall circuit diagram	
	TOTAL OF THE USE DEPOSITING FOR CHALITY	
6.	CODING SYSTEM OF FAILURE REPORTING FOR QUALITY	

ADDRESSES FOR SALES AND SERVICE

7.

INHALTSVERZEICHNIS

TABLE DES MATIERES

1.	ALLGEMEINES	17	1.	GENERALITIES	29
1.1.	Einleitung	17	1.1.	Introduction	29
1.2.	Technische Daten	17	1.2.	Caractéristiques techniques	29
1.3.	Zubehör	21	1.3.	Accessoires	33
1.4.	Funktionsprinzip	22	1.4.	Principe de fonctionnement	33
2.	VORBEREITUNGSANWEISUNGEN	23	2.	INSTRUCTIONS POUR L'INSTALLATION	35
2.1.	Wareneingangskontrolle	23	2.1.	Inspèction initiale	35
2.2.	Sicherheitsanweisungen	23	2.2.	Consignes de sécurité	35
2.3.	Netzspannungseinstellung und Sicherungen	24	2.3.	Adaptation à la tension secteur, fusibles	35
2.4.	Betriebslage des Gerätes	24	2.4.	Position de fonctionnement de l'appareil	36
2.5.	Öffnen des Gehäuses	24	2.5.	Démontage de l'appareil	36
2.6.	Funkentstörung	24	2.6.	Suppression du rayonnement	36
3.	BETRIEBSANLEITUNG	25	3.	MISE EN SERVICE	37
3.1.	Bedienelemente und Anschlüsse	25	3.1.	Commandes et douilles	37
3.2.	Bedienung	26	3.2.	Fonctionnement	38

5. BILDVERZEICHNIS

- 1 Blockschaltbild
- 2 Frontansicht
- 3 Rückansicht

5. RAPPEL DES FIGURES

- 1 Schéma synoptique
- 2 Face avant
- 3 Face arrière

Operating manual

ì 1

GENERAL

1.1. INTRODUCTION

The PM 5132 function generator is an instrument designed for applications extending from the educational to the general purpose area.

It produces sinewave, triangular and squarewave output signals as well as positive and negative pulses, the frequencies of which are adjustable in seven linear sub-ranges from 0.1 Hz to 2 MHz. The frequency vernier allows the frequency setting to be varied from -5% to +5%.

The output voltage is continuously adjustable from 3 Vpp to 30 Vpp and can be attenuated in steps of 10 dB to 60 dB.

A continuously adjustable d.c. output voltage can be selected separately or whenever used as d.c. offset voltage added to the selected output signal.

The internal resistance of the generator can be selected by means of a pushbutton between 50 Ω and 600 Ω .

The duty cycle can be set to a fixed value of $50\,\%$ or is continuously adjustable between $10\,\%$ and $90\,\%$.

The generator frequency is controllable in each of the seven sub-ranges (internal sweep). The sweep period is adjustable within a range of 0.05 s to 100 s. Moreover external sweep and frequency modulation can be performed.

For TTL applications a separate output is available.

The ergonomic design of controls and sockets serves for convenient operating of the instrument.

1.2. TECHNICAL DATA

Safety characteristics

This apparatus has been designed and tested in accordance with Safety Class I requirements of IEC Publication 348, Safety Requirements for Electronic Measuring Apparatus, and has been supplied in a safe condition. This manual contains some information and warnings which must be followed by the user to ensure safe operation and to retain the apparatus in a safe condition.

Performance characteristics, specifications

Properties expressed in numerical values with stated tolerance are guaranteed by the manufacturer. Specified non-tolerance numerical values indicate those that could be nominally expected from the mean of a range of identical instruments.

This specification is valid after the instrument has warmed up for 30 minutes (reference temperature 23 °C).

If not stated otherwise, relative or absolute tolerances relate to the set value.

1.2.1. Frequency

Frequency range		0.1	Hz	-	2	MHz
Selected range I		0.1	Hz	-	2	Hz
١	I	1	Hz	-	20	Hz
1	H	10	Hz	-	200	Hz
,	V	100	Hz	-	2	kHz
,	v	1	kHz	-	20	kHz
,	VI.	10	kHz	-	200	kHz
•	VII.	100	kHz	-	2	MHz

1.2.2.

linear Characteristic Adjustments - seven range-pushbuttons - dial with linear scale - knob for fine-adjustment Frequency indication - linear scale on the dial ± 2 % of dial setting $\pm 0,2$ % of full scale Setting error Vernier frequency adjustment ±5 % of the dial setting <0,1 %/K ranges II - VI Temperature coefficient <0,15 %/K ranges I - VII < 0.10 % ranges I - VI Short term drift within 15 min < 0.50 % range VII < 0.25 % ranges I - VI Long term drift within 7 hours < 1.50 % range VII Signal output Connection **BNC** socket 50 $\Omega/600 \Omega$, selected by pushbutton Impedance short-circuit proof Load capability Wave forms sinewaye triangular wave □ squarewave n-positive pulse negative pulse DC d.c. voltage without a.c. adjustable from 10 % to 90 % when pushbutton Duty cycle DUTY CYCLE is pressed (restricted use in range VII) Open-circuit voltage 3 Vpp to 30 Vpp, continuously adjustable - setting range 1,5 Vpp to 15 Vpp for pulses -maximum value ±15 V DC (offset) voltage - button PUSH FOR ZERO pulled, open circuit voltage -10 to +10 V, continuously adjustable - button PUSH FOR ZERO or WAVE FORM button DC pressed < 50 mV Attenuation - continuous 0 to 20 dB (output voltage adjustable from 3 Vpp to 30 Vpp) - fixed 0 to 60 dB in steps of 10 dB Distortion (sinewave) < 0,5 % in ranges I to V < 1 % in range VI < 3 % in range VII Linearity (triangular wave) > 99 %

risetime, falltime (squarewave)

overshoot and ringing

(squarewave)

Amplitude response (sinewave, reference value = 1 k Hz;) < 75 ns. (at max. amplitude, load 50 Ω)

< 2 % (at max. amplitude, load 50 Ω)

< 0.1 dB in ranges I to V

< 0.3 dB up to 1 MHz < 1 dB in ranges I to VII

(at max. amplitude,load 50 Ω , attenuation 0 dB).

1.2.3. TTL-output

Connection

Duty cycle, frequency

Fan out

BNC socket at the rear side

identical with main output


20 TTL inputs

1.2.4. Frequency control

1.2.4.1. Internal sweep

Sweep modes

- single sweep started by a pulse (SWEEP TRIG INPUT)
- single sweep started by pressing the pushbutton TRIG
- continuous sweep



Sweep characteristic

Start frequency

Stop frequency

Sweep period

Sweep control elements

linear

continuously adjustable within the respective frequency range after the pushbutton STD BY / CONT is set to STD BY.

identical to the dial adjustment

50 ms to 100 s, adjustable in three ranges:

- 50 ms to 1 s
- -0,5 s to 10 s
- −5 s to 100 s
- pushbutton to start a single sweep
- pushbutton to stop the running sweep immediately
- pushbutton for resetting to f $_{\mathsf{STAR}}$ during single sweep.

SWEEP TRIG INPUT

- Connection

BNC socket at the rear side

- Impedance

> 10 k Ω

- Max. input voltage

±15 V

- Release criterion

negative going edge of a pulse (from High to Low);

e.g. TTL signal with slew rate > 0.5 V/ms

SWEEP VOLTAGE OUTPUT

- Connection

BNC socket at the rear side

- Impedance

10 k Ω

- Output voltage

0 V to 5 V (f_{START} ... f_{STOP})

PEN LIFT

- Connection

BNC socket at the rear side

electronic switch, open during flyback,

closed during the sweep.

max. current: 200 mA — electronic switch

output voltage: < 0,7 V I closed

output voltage: +22 V ---- electronic switch

internal resistance: 100 kΩ ¹ open

1.2.4.2. External sweep or frequency modulation (SWEEP INPUT)

- Connection

BNC socket

- Voltage vs. frequency characteristic

linear

Impedance

 $47\,k\,\Omega$

- Max. sweep range

2 1/2 decades

- Max. sweep frequency

15 k Hz

- Sensitivity

1 V / 0,2 $\rm f_{max}$, where $\rm f_{max}$ represents the upper limit

of the corresponding subrange.

1.2.5. Power supply

AC mains

- Reference value

220 V

- Nominal vaues

110 V / 128 V / 220 V / 238 V, selectable by solder

links

- Nominal operating range

±10 % of selected nominal value

Operating limits

±10 % of selected nominal value

- frequency range of operation

47,5 - 105 Hz

Power consumption

25 W

1.2.6. Environmental conditions

Ambient temperature

- Reference value $+23^{\circ}$ C $\pm 1^{\circ}$ C - Nominal operating range $+5^{\circ}$ C to $+40^{\circ}$ C

- Limits for storage and transportation -40° C to +70° C

Relative humidity

Reference range
Nominal operating range
Limits for storage and transportation
45 % to 75 %
20 % to 80 %
0 % to 90 %

Air pressure

- Reference value 1013 mbar (= 760 mmHg)

- Nominal working range 800 - 1066 mbar (up to 2200 m height)

Air speed

Reference valueNominal working range0 to 0,2 m/s0 to 0,5 m/s

Operating position normally upright on feet or with handle fold down

Warm-up time 30 min.

1.2.7. Cabinet

Protection type

(see DIN 40 050) IP 20

Protection class

(see IEC 348) class I, protective conductor

Overall dimensions

 - height
 140 mm

 - width
 310 mm

 - depth
 330 mm

 Weight
 approx. 5 kg

1.3. ACCESSORIES

1.3.1. Standard Instruction manual,

fuse and voltage labels

1.3.2. Optional PM 9585: 50 Ω termination 1 W

PM 9581: 50 Ω termination 3 W

PM 9075: Coaxial connection cable BNC-BNC

1.4. OPERATING PRINCIPLE (see Fig. 1., block diagram)

The main oscillator of the function generator comprises the switched integrator and the peak detector (comparator). The controlled current source of the d.c. control section generates the charging current for the integrator. At the integrator output a linear voltage ramp is fed to the peak detector. When reaching the reference voltage the detector reverses the charging current of the integrator resulting in integration in the opposite direction.

Integration down is performed until reaching the negative reference level of the peak detector, which again reverses the current of the switched integrator. As both reference levels are symmetrical with respect to earth, a zero symmetrical triangular wave is generated at the output of the integrator. The duty cycle of this wave is 1:1 and can be adjusted continuously between 1:9 and 9:1 when push-button DUTY CYCLE is pressed.

The output current of the controlled current source depends on the positions of the frequency dial and the FREQUENCY OFFSET control. The resulting frequency of the oscillator is determined by this current and -in addition- by the integrating capacitor in the switched integrator. Different capacitors are switched in by the FREQUENCY Hz pushbuttons.

External sweep and frequency modulation can be performed via the SWEEP INPUT socket. The voltage to frequency relationship is linear.

Internal frequency control is effected by the sweep oscillator, a continuous sweep is started by setting the pushbutton STD BY/CONT to position CONT. Starting a single sweep can be done by pressing the pushbutton TRIG. Moreover it is possible to start a single sweep by means of an external trigger pulse connected to the SWEEP TRIG INPUT socket.

The sweep range is adjustable by means of the knob START FREQ and the dial which represents the stop frequency. They are adjustable independently; the start frequency may be set greater or less than the stop frequency.

Moreover the outputs PEN LIFT and SWEEPOUTare controlled by the sweep oscillator. The PEN LIFT socket represents the output of an electronic switch being closed during the sweep. The SWEEP OUT socket gives the instantaneous sweep voltage corresponding to a distinct frequency of the oscillator: 0 V = start frequency, +5 V = stop frequency f_{STOP} . The voltage to frequency relationship is linear.

The WAVE FORM switch allows the following wave forms to be selected: a triangular wave from the integrator, a sine wave formed by the sine shaper circuitry from the triangular wave, a square wave, negative and positive pulses from the peak detector.

The selected signal is fed to the OUTPUT socket via amplifier and attenuator. The internal resistance can be switched to 600 Ω or to 50 Ω .

By means of the DC OFFSET control a d.c. voltage can be added to the signal, activated by pulling the PUSH FOR ZERO switch/turn-knob. If d.c. voltage only is requested, the a.c. part of the output signal can be switched off by pressing the DC pushbutton of the wave form selector.

A squarewave signal of the peak detector is fed to the socket TTL OUT via the TTL buffer. The stabilized power supply provides the d.c. voltages for the circuitries.

2. INSTALLATION INSTRUCTIONS

2.1. INITIAL INSPECTION

Check the contents of the shipment for completeness and note whether any damage has occurred during transport. If the contents are incomplete, or there is damage, a claim should be filed with the carrier immediately, and the Philips Sales or Service organisation should be notified in order to facilitate the repair or replacement of the instrument.

2.2. SAFETY INSTRUCTIONS

Upon delivery from the factory the instrument complies with the required safety regulations, see para.

1.2. To maintain this condition and to ensure safe operation, the instructions below must carefully be followed.

2.2.1. Maintenance and repair

Failure and excessive stress:

If the instrument is suspected of being unsafe, take it out of operation permanently.

This is the case when the instrument

- shows physical damage
- does not function anymore
- is stressed beyond the tolerable limits (e.g. during storage and transportation)

Dismantling the instrument: When removing covers or other parts by means of tools, live parts or terminals could be exposed. Before opening the instrument, disconnect it from all power sources.

If the open live instrument needs calibration, maintenance or repair, it must be performed only by trained personnel being aware of the risks. After disconnection from all power sources, the capacitors in the instrument may remain charged for some seconds.

2.2.2. Earthing (grounding)

Before any other connection is made the instrument shall be connected to a protective earth conductor via the three-core mains cable. The mains plug shall be inserted only into a socket outlet provided with protective earth contact. The protective action shall not be negated by the use of an extension cord without protective conductor.

The external contacts of the BNC sockets must not be used to connect a protective conductor.

WARNING: Any interruption of the protective conductor inside or outside the instrument, or disconnection of the protective earth terminal, is likely to make the instrument dangerous. Intentional interruption is prohibited.

The circuit earth potential applied to the external contacts of the BNC sockets is connected to the cabinet. The external contacts of the BNC sockets must not be used to connect a protective conductor.

2.2.3. Connections

The circuit earth potential is applied to the external contacts of the BNC sockets and is connected to the cabinet by means of parallel-connected capacitor and resistor. By this means hum loops are avoided and a clear HF earthing is obtained.

If the circuit earth potential in a measurement set-up is different from the protective earth potential, it must be noticed,

- that the BNC sockets can be touched and that it must not be live, see the safety regulations on the subject (VDE 0411),
- that all sockets marked with the sign \(\precedut \) are internally interconnected.

2.2.4. Mains voltage setting and fuses

Before inserting the mains plug into the mains socket, make sure that the instrument is set to the local mains voltage.

The instrument shall be set to the local mains voltage only by a qualified person who is aware of the bazard involved.

WARNING: If the mains plug has to be adapted to the local situation, such adaption should be done by a qualified person only.

Make sure that only fuses of the required current rating, and of the specified type, are used for renewal. The use of repaired fuses, and/or the short-circuiting of fuse holders, are prohibited. The fuse shall be renewed only by a qualified person who is aware of the hazard involved.


WARNING: The instrument shall be disconnected from all voltage sources when a fuse is to be renewed, or when the instrument is to be adapted to a different mains voltage.

2.3. MAINS VOLTAGE SETTING AND FUSES

The safety instructions in chapter 2.2.4. must be followed.

On delivery from the factory the instrument is set to 220 V (PM 5132) resp. 110 V (PM 5132 M). If the instrument is to be used on a different supply voltage proceed as follows:

- Unplug the mains connector
- Fold up the handle to the top.
 For this push the buttons of the handle.
- Loosen the central screw at the rear
- Dismantle the cabinet
- Change the solder links according to the connection diagram



- If necessary exchange the supplied fuse and mains voltage label
- Close the instrument

2.4. OPERATING POSITION OF THE INSTRUMENT

The instrument may be used in the positions indicated in clause 1.2. 6. With the handle folded down, the instrument may be used in a sloping position; for this push the buttons of the handle. The characteristics mentioned in Section 1.2. are guaranteed for the specified positions.

Ensure that the ventilation holes in the cover are free of obstruction.

Do not position the instrument on any surface which produces or radiates heat, or in direct sunlight.

2.5. DISMANTLING THE INSTRUMENT

- Unplug the mains connector
- Fold up the handle to the top. For this push the buttons of the handle
- Loosen the central screw at the rear
- Dismantle the cabinet

2.6. RADIO INTERFERENCE SUPPRESSION

Radio interference of the instrument is suppressed and checked carefully. In connection with deficient suppressed base units and further units radio interference can be generated, which have to be suppressed by means of additional activities.

3. OPERATING INSTRUCTIONS

3.1. CONTROLS AND SOCKETS (FIG. 2, 3)

Legend	Position	Function
POWER ON OFF	251	mains switch; white dot for ON position
WAVE FORM	253/9 to 253/4	pushbuttons for the required waveform: sinewave, triangular-, square-wave, positive or negative pulses, DC without ac portion
FREQUENCY Hz x.1 x1 x10 x100 x1K x10K x100K	253/16 to 253/10	pushbuttons for selecting the frequency range 0,1 - 2 Hz, 1 Hz - 20 Hz, 10 Hz - 200 Hz, 100 Hz - 2 k Hz, 1 kHz - 20 kHz, 10 kHz - 200 kHz, 100 kHz - 2 MHz
1 20 (Linear scale)	648	dial for continuous coarse frequency adjustment
FREQ OFFSET	647	knob for continuous fine frequency adjustment
DUTY CYCLE	252/5	pushbutton to switch on the duty cycle adjustment
DUTY CYCLE	673	knob for duty cycle adjustment
ATTENUATION 10 dB 20 dB 30 dB	253/3 to 253/1	pushbuttons for setting the fixed attenuation; 40 dB to 60 dB by combination of pushbuttons
AMPLITUDE 3 Vpp to 30 Vpp	821	knob for continuous amplitude adjustment of the output signal.
DC OFFSET -10 V to +10 V	822 .	knob for continuous d.c. voltage adjustment
PUSH FOR ZERO	822	pull-switch for switching in the d.c. voltage
OUTPUT	875	BNC socket for the output signal
TTL OUT	874	BNC socket for TTL output signal (at the rear side)
600 Ω/50 Ω		pushbutton for selecting the impedance of the main output.
PEN LIFT	872	control signal for pen writing in case of plotting (BNC socket at the rear side).
SWEEP INPUT	873	BNC socket at the rear side; input of external sweep voltage
SWEEP OUT	871	BNC socket at the rear side; output of the sweep voltage during internal sweep
START FREQ CHECK IN STD BY	646	knob for adjusting the start-frequency

TRIG	252/4	pushbutton for starting a single sweep
■ STD BY	252/3	STD BY - for adjusting the start frequency and for single sweep
CONT		CONT - continuous sweep
SWEEP PERIOD s	256	range selector for the sweep period
PERIOD	601	knob for adjusting the sweep period
HOLD	252/1	pushbutton for stopping the sweep immediately
RESET	252/2	pushbutton for resetting during single sweep to f_{START} .
SWEEP TRIG INPUT	870	BNC socket at the rear side — to start a sweep by means of an external signal.

3.2. OPERATION

3.2.1. Setting the voltage at socket OUTPUT

By means of the control AMPLITUDE, the amplitude of the output signal is continuously variable. Released button DC and pulled button PUSH FOR ZERO enables a continuously adjustable positive or negative d.c. voltage to be added to the output signal.

When pressing the button DC, the a.c. part of the output signal is switched off and the d.c. voltage only is fed to the output.

With step attenuator ATTENUATION, the output signal and the DC offset can be attenuated in steps of 10 dB up to 60 dB. For 10 dB to 30 dB separate pushbuttons are available. Attenuations of 40 and 50 dB are selected by combined pushbutton actions.

For 60 dB all three buttons have to be pressed.

By means of the pushbutton 600 $\Omega/50$ Ω the internal resistance can be switched to 600 Ω or 50 Ω .

Note:

The output amplifier could be overdriven due to adding signal and DC offset voltage. To avoid limiting, the peak value of the open-circuit output voltage must not exceed ± 15 V (step attenuator set to 0 dB).

3.2.2. Setting the frequency

The PM 5132 offers three control elements for setting the frequency:

- a range selector FREQUENCY Hz with seven ranges
- a dial with linear scale
- vernier FREQ OFFSET control

The output frequency is represented by the scale reading multiplied by the factor of the range selector FREQUENCY Hz. In addition the frequency deviation set of the FREQ OFFSET control must be accounted for.

3.2.3. Setting the wave form

The required wave form is selected by pressing the according pushbutton of the wave form selector. If the duty cycle shall be set to any value different to 50 %, the pushbutton DUTY CYCLE must be pressed fo adjusting the duty cycle within 10 % and 90 % by means of the knob DUTY CYCLE.

At the TTL OUT socket # TTL signal is available, the frequency and duty cycle of which correspond to the signal at the OUTPUT socket.

3.2.4. Setting the internal sweep

The internal sweep action of the PM 5132 can be performed within one subrange.

The value of the dial represents the stop frequency, the start frequency is set according to the following

- pushbutton STD BY / CONT must be set to STD BY
- range selector SWEEP PERIOD s must be switched from OFF to the required range
- adjust the sweep period with the knob PERIOD
- adjust the start frequency with knob START FREQ. This frequency can now be measured at the socket OUTPUT.

The start frequency may be set greater or less than the stop frequency, a sweep can be performed in both directions, up and down.

Starting a single sweep is done by pressing the pushbutton TRIG once, the continuous sweep is realized by setting the pushbutton STD BY/CONT to CONT.


In position STD BY a single sweep can be started by a triggering pulse at the SWEEP TRIG INPUT socket. Starting criterion is a negative going edge e.g. of a TTL signal.

The SWEEP OUT socket at the rear side provides the instantaneous sweep voltage corresponding to a distinct signal frequency for controlling an oscilloscope or x - y plotter. The voltage to frequency relationship is linear, a sweep-voltage of 0 V represents f_{STABT} , a sweep voltage of +5 V f_{STOP} .

Further control elements for manual controlling are the HOLD pushbutton to stop the sweep immediately and RESET for resetting to f_{STABT} during single sweep.


Note: Before executing an internal sweep, disconnect all cables from the SWEEP INPUT socket.

This input is not switched off during an internal sweep.



3.2.5. External sweep and frequency modulation

The signal frequency of the generator can be controlled by connecting an external voltage to SWEEP INPUT. In order to avoid superpositions by the internal sweeps, switch the range selector SWEEP PERIOD s to OFF. The desired signal frequency (= basic frequency) must then be adjusted by means of the dial and the range selector FREQENCY Hz. By feeding a dc or ac voltage to the sweep oscillator via SWEEP INPUT, the signal frequency is varied proportional to this voltage (the relationship between voltage and frequency is linear). The maximum variation should not exceed 2 1/2 decades, where the upper limit of the sub-range is the upper limit for the sweep range too. Otherwise you would get signal distortions because of overdriving the controlled current source.



External sweep

Sweep range depending on the sweep voltage Uc. Example for subrange V (1 $\,\mathrm{KHz}$ - 20 $\,\mathrm{KHz}$)

Service part

4.1. CIRCUIT DESCRIPTION PM 5132

4.1.1. Sweep oscillator

In sweep mode the sweep oscillator generates sawtooth voltage of 10 Vpp to control the main oscillator. The sweep oscillator mainly consists of the integrating operational amplifier 401 with charging capacitor 501 to 504, comparator 402 and regulator 301/302. The potmeter PERIOD 601 determines the charging current which is fed via resistor 616 and SK 252/1a to the integrator.

In sweep mode a positive voltage is applied to the input of the integrator, resulting in a slowly falling ramp at its output. The reference level of the comparator is set to 0 V and transistor 301 is turned off. As soon as the integrator voltage has reached this value the comparator turns over. So during fly-back the output voltage of the comparator is fed via conducting diode 421, resistor 615 and additionally via potmeter PERIOD and resistor 616 to the input of the integrator. So the integration capacitor is quickly discharged. When the integrator voltage has reached +10 V the comparator turns over and a new cycle starts.

In STAND BY mode the cycle just mentioned is interrupted after fly-back. The positive reference level of the comparator 402 is set to +11 V by opening the parallel path SK 252/3a and resistor 622. So the comparator cannot turn over. Regulator 301/302 is activated via resistor 606 to hold the integrator at the positive peak value. Addionally the integrator voltage is fed via divider 612, 614 to the regulator. The amplitude is limited to 4 Vpp by clamping the resistor 613 and diodes 422/423 to the base of transistor 302 to avoid too high emitter base reverse voltage of 301.


When pressing pushbutton TRIG a negative pulse is applied to the reference level of the comparator 402 via resistor 641/642 and the comparator is switched over from its STAND BY position; so one cycle starts with the slowly falling ramp of the integrator. Overmore triggering is possible by external TTL-signal on the falling edge via SWEEP TRIGGER INPUT.

When pressing pushbutton RESET a positive pulse is fed to the inverting input of the comparator. So fly-back is performed.

In HOLD mode current to the integrator is interrupted by SK 252/1a and the integrator remains at its instantaneous value.

The signal of the comparator is additionally applied via diode 427 to inverting switching transistor 306 to control the PEN LIFT output. During the sweep the PEN LIFT output is set to ground. In fly-back and stand-by position electronic switch 306 opens and +22 V is routed to the output.

In sweep mode the SWEEP OUTPUT is controlled by the integrator voltage, inverted via amplifier 401/2 and halved by divider 629/631 (see diagram chapter 3.2.4.).



4.1.2. Control section

The general task of the control section is to generate frequency proportional charging currents for the main oscillator. Input signals of this circuitry are the voltages at the potentiometers for the start and stop frequency, the output voltage of the sweep oscillator and the external sweep or modulation signal.

In normal mode (SWEEP OFF) the input of the amplifier 401.2 is set to ground; so the start potmeter is not active. As the output of the amplifier is U = 2 Uref-Usweep, where Uref = +5 V, +10 V are applied to the potmeter 648 for frequency setting. The voltages at the wipers are converted to currents via resistors 651, 652 and are fed to the summing amplifier 404. The output signal is converted into control voltages by current determining resistors 656 to 659, 662 and 664 for the 4 different frequency ranges. For the lowest range of the main oscillator the resistor is reduced by the factor 10 and the capacitors in the quadruple switch are not increased; so the size of the capacitors is limited.

Possibly scattered-in spikes via leads of the control section are suppressed via capacitor 531.

In SWEEP mode and stand-by operation the output of +10 V of the integrator of the sweep oscillator is connected to inverting amplifier 401/2 and to the start potentiometer 646. So 0 V is applied to the SWEEP OUTPUT socket; the (stop) FREQUENCY potmeter 648 is inactive. By this independent setting of the start and stop frequency for the sweep is achieved.

During the sweep the voltages at the two potmeters are decreasing and increasing in mutual direction until at the end of the sweep frequency potmeter 648 only is active as in normal mode.

When DUTY CYCLE is not pressed potmeter 673 is not active. Control voltage is 1:1 passed via impedance converter OP 405/transistor 312 to the negative current source, comprising amplifier 407 and transistor 314 and to the auxiliary current source 406. This negative current source is the reference for the positive current source the control of which is achieved via the symmetry control. The auxiliary current source 406 generates a current through resistor 667, which is fed to resistor 654. The resulting voltage generates the basic current for the positive current source, comprising amplifier 408 and transistor 313.

The symmetry control 315 to 318 was introduced to equalize for any inaccuracies caused by distortion and unbalance of the positive and negative current source and of the 4 transistors of the integrator. On the virtual earth side of the selected charging capacitor, the capacitor 515 will be charged by any unbalance, because it is inserted between charging capacitor and earth. This unbalance provides a control voltage for FET 318. The differential stage operates as a voltage/current converter, the output either putting current into the current source or taking out of it depending on the state of the positive current control. By this means, and by transistors with high current gain (i.e. negligible leakage current) in the quadruple switch circuit, any unbalance of the two loading currents is equalized, so reducing time symmetry errors of the signal.

When the DUTY CYCLE button is pressed the ratio of the two charging currents can be varied by means of the potentiometer DUTY CYCLE. So the duty cycle of the signal generated by the main oscillator is varied. At amplifier 405 divider 668/669/671 now is connected to the negative reference potential. So only a portion of the output voltage of the amplifier is fed back resulting in increased gain of 5. Furthermore resistors 672, 674 are switched to the negative reference potential, so that the DUTY CYCLE potmeter is active.

Variation of the control voltage at potentiometer 673 results in load current variation from 0.2 to 5; so the duty cycle is adjustable between 10 % and 90 %.

The period time in duty cycle mode is adjusted by 669 to the same value as in normal mode. Symmetric arrangement of the duty cycle circuitry guarantees equal period time of the main oscillator, when the DUTY CYCLE potmeter is turned.

Furthermore in duty cycle mode the symmetry control has to be switched off which is done by connecting the virtual zero of the quadruple switch to ground and by separating the output of the symmetry control from the positive current source.

4.1.3. Main oscillator

The wave form generator operates on the relaxation oscillator principle whereby the selected range capacitor first charged linearly in one direction and then charged linearly in the reverse direction to produce a triangular wave form. The frequency is determined by the selected range capacitor and the charging current, which can be influenced by a control voltage input derived from the frequency dial setting and, if required, by a sweep voltage input.

Switchover of the charging current is achieved by a two-level detector or comparator 328/329 which produces a reference signal when the selected integration capacitor has charged to a predetermined level. This reference signal is fed back to reverse the charge through the capacitor by means of the quadruple switching circuit.

The circuit operates as follows:

Assume that point a is positive. The integration capacitor will charge via transistor 321 and 320 (319 and 322 are switched off). At a predetermined level, transistor 328 of the comparator will switch on and open the current switch 332. Current will flow from the current source transistor 330 via 332 to switch off the transistor 329 of the comparator. Resulting negative reference voltage of the comparator at resistors 702/703 is routed to point a. The quadruple switch circuit will switch over. Integration capacitor will now be charged in the reverse direction via transistor 319 and 322. At a negative predetermined level, transistor 328 will switch off and also current switch 332. Current switch 331 will turn on because transistor 329 is active. So point a has positive voltage again—and a new cycle starts. In this way, the triangular wave form is generated, the frequency of which depends on the charging capacitor and the charging current. The square wave is produced by 332.

The basic part of the oscillator is the quadruple switch circuit with the frequency determining capacitors 516 - 523, selected by the front-panel pushbuttons FREQUENCY Hz. Under the control of the squarewave signal at point a, at each half-cycle two diagonal opposite transistors open while the other two close (i. e. 321, 320 open, 319, 322 close, vice versa). In this way the direction of the charging current is changed. To produce a triangular waveform that is time-symmetrical with respect to earth, one side of the charging capacitor is connected to virtual earth at the gate input of FET 318. In this way, the triangular signal is taken from the active side of the charging capacitor with respect to the 0 V line.

A high impedance FET buffer 323 avoids charging current variations occurring on the active side of the charging capacitor. Emitter follower 325/326 then connects the signal to the comparator 328/329.


To compensate for non-linearity in the frequency response of the highest range due to circuit delays in the comparator and the quadruple switch the lead 527 is inserted between oscillator and comparator.

The comparator 328/329 compares the instantaneous value of the capacitor charging potential against the reference level at the base 329 which is ± 2 V square wave symmetrical. When the circuit attains one of the reference levels, it actuates the quadruple switch. At the same time, the comparator reference level is switched to the opposite polarity.

On the reference side of the comparator 2 current sources generate voltages via resistors 702, 703. The current of the first source, comprising transistor 330/diode 447, has twice the value of the second source, comprising resistors 704/705. According to the state of transistor 331 either the difference of the two sources results in a +2 V reference level or the second source alone generates -2 V reference level.

When square wave form is chosen, all switches are open, see figure below. The current of the current switch 332 (active € 20 mA) flows via resistor R1. Negative current is added via resistor R2, resulting in zero-symmetrical square wave signal.

Switching in positive or negative pulses resistor R3 is added by S1. For R1 = R3 the square wave amplitude is halved. When R4 is activated by switch S2 or R5 by S3 positive or negative current is added, resulting in a square wave signal between 0 V and its positive value or between 0 V and its negative value. The total square wave signal is routed via a complementary emitter-coupled buffer stage 333/334 to the WAVE FORM switch 253. The signals are inverted by the power amplifier.



S 1 to S3 in position square wave

R1 = 714, 715, 716 + 710

R2 = 711

R3 = 717

S1 = 253/5b or 253/6a

S2 = 253/5a for negative pulses at the

S3 = 253/6b for positive pulses \(\) OUTPUT

The TTL output is taken from resistor 705 and routed via differential amplifier 335/336 and output switch 337. The output is taken directly from the emitter of 337 to the TTL OUTPUT socket at the rear side.

4.1.4. Sine shaper

The triangular signal of the oscillator is fed via WAVE FORM switch to the sine shaper circuit 412. The conversion from triangle to sine wave, with a distortion factor of 0.2 % typical, is achieved by non-linear wave form shaping at a diode-resistor network in the IC. The peaks of the triangular wave form are suppressed by a special compensating circuit in the integrated circuit.

Associated preset controls for shaping are potentiometers 808 and 812. The output on pin 1 is routed via a low-pass filter which serves to reduce the harmonic content of the sine wave at the highest frequencies. Preset 814 provides a control adjustment for the amplitude response. The output d.c. offset is preset to zero by 816. The emitter-follower 350/351 is used for impedance matching. The output is fed via the wave form selector switch and front-panel AMPLITUDE control 821 to the power amplifier.

4.1.5. Power amplifier

The power amplifier mainly consists of a voltage amplifier, output stage and the attenuator. Via the AMPLITUDE potmeter the level of the selected signal is applied to the voltage amplifier 352 - 355, which operates as a complementary cascode stage in pull-push arrangement to drive the power output. The output stage comprises 356/358 and 357/359 and operates with complementary darlingtons. Feedback is done via resistor 831 with parallel capacitor 556 to the input of the voltage amplifier, resulting in overall gain of ca. 15 dB.

The DC OFFSET is controlled by potmeter 822 via resistor 826 to the input of the voltage amplifier and can be set up to $\pm 10~V$; see also note in chapter 3.2.1. Resistors 846/847, which determine the output impedance, feed the three-stage-attenuator. The front panel ATTENUATION pushbutton 253/1/2/3 permits selection of 10 dB, 20 dB and 30 dB by switching resistors 849 to 855 as L-section attenuator. For 600 Ω output impedance the signal is connected via resistor 857 to the OUTPUT socket.

4.1.6. Power supply

The required two power supplies of ± 22 V are realized by means of four terminal adjustable voltage regulator 410 and 411. Positive voltage is adjusted by 781 and negative voltage by means of 784 to an accuracy of ± 0.1 V.

4.2. ACCESS TO PARTS

Before dismantling the instrument, the safety regulations in accordance with para. 2.2. must be strictly observed.

4.2.1. Cabinet, see 2.5.

4.2.2. Knobs

- Remove the cap from the knob.
- Unscrew the nut and remove the knob.
- When replacing the knob, ensure that the white mark is correctly aligned with the text plate markings.

4.2.3. Text plate

- Remove the cabinet, see 2.5.
- Remove the turn-knobs, see 4.2.2.
- Remove the dial.
- Remove the plastic cover of the mains switch.
- The text plate can now be removed.
 - Be careful:

The textplate is fitted to the frontplate by double sided adhesive tape.

4.2.4. Pushbutton unit


Replacing a pushbutton lever.

The single pushbutton lever can be replaced from the front.

- Push the spring towards the pushbuttons.
- Remove the wire strap and/or lift the plastic reed between the contacts
- Carefully tear the pushbutton lever out of the pushbutton.

Replacing a switch of the pushbutton unit

- Straighten the 4 retaining lugs of the relevant switches as shown in the figure below.
- Break the body of the relevant switch by means of a pair of pliers and remove the pieces. The soldering pins are then accessible.
- Remove the soldering pins and clean the holes in the printed circuit board (e.g. with a suction soldering iron).
- Bend the 4 retaining lugs back to their original positions.
- Solder the new switch on to the printed circuit board.



4.3. CHECK AND ADJUSTMENT

4.3.1. General

- The limits mentioned in this paragraph are valid only for a newly adjusted instrument and therefore might deviate from the values as stated in paragraph 1.2. "Technical Data".
- Adjustment of the instrument is only permitted after a warm-up time of at least 30 minutes at an ambient temperature of (+23 ±3)° C and when connected to a mains voltage of 220 V ±5 %. The cabinet must be closed.
- The printed circuit board is mounted overhead. Nevertheless all trimming potentiometers and capacitors are accessible from the top. For adequate temperature stability during adjustment, the cabinet should be removed only for a short time and so far that the required adjusting element is just accessible.
- If not explicitely stated otherwise, the voltage potentials refer to the relevant contact measured against circuit earth (⊥_____).
- The following abbreviations are used for setting and measuring instruments:

X = Button pressed

= Button not pressed/unlocked

o = Button only tipped

D.V.M. (d.c.) = Digital voltmeter for DC measurements e.g. PM 2534
OSC. = Oscilloscope e.g. PM 3055
C = Counter e.g. PM 6665
Fg = Function generator e.g. PM 5132
DA = Distortion analyzer e.g. HP 333A

50 Ω = 50 Ω terminating resistor e.g. PM 9585

4.3.2. Preparations

- All trimming potentiometers and capacitors in mid-position; (only for complete new adjustment).
- Turn dial potentiometer to extreme counter clockwise position.
 Position dial such that the 0,4 graduation on the dial is positioned to the text plate mark.
- Solder joints A to K must be closed. To be opened for failure detection only.
- Terminate the OUTPUT by a 50 Ohm resistor.

4.3.3. General functional test

- Actuate all buttons one after the other for rough functional test of the generator by means of an oscilloscope connected to the OUTPUT socket.
- Control the TTL output.
- Roughly control the output voltage at the SWP VOLTAGE socket during the internal sweep.

4.4. SAFETY INSPECTION AND TESTS AFTER REPAIR AND MAINTENANCE IN THE PRIMARY CIRCUIT

4.4.1. General directives

- Take care that creepage distance and clearances have not been reduced
- Before soldering, wires:
 should be bent through the holes of solder tags, or wrapped round the tag in the form of an open
 U, or, wiring ridigity shall be maintained by cable clamps or cable lacing.
- Replace all insulating guards and -plates.

4.4.2. Safety components

Components in the primary circuit may only be renewed by components selected by Philips, see also chapter 4.5.1.

4.4.3. Checking the protective earth connection

The correct connection and condition is checked by visual control and by measuring the resistance between the protective-lead connection at the plug and the cabinet/frame. The resistance shall not be more than 0.5 Ω . During measurement the mains cable should be moved. Resistance variations indicate a defect.

4.4.4. Checking the insulation resistance

Measure the insulation resistance at U = 500 Vdc between the mains connections and the protective lead connections. For this purpose set the mains switch to ON. The insulation resistance shall not be less than 2 $M\Omega$.

Note:

2 M Ω is a minimum requirement at 40 $^{\rm o}$ C and 95 % relative humidity. Under normal conditions the insulation resistance should be much higher (10 to 20 M Ω).

4.5. SPARE PARTS

4.5.1. General

Standard Parts

Electrical and mechanical parts replacement can be obtained through your local Philips organisation or representative. However, many of the standard electronic components can be obtained from other local suppliers. Before purchasing or ordering replacement parts, check the parts list for value, tolerance, rating and description.

NOTE:

Physical size and shape of a component may affect instrument performance, particularly at high frequencies. Always use direct-replacement components, unless it is known that a substitute will not degrade instrument performance.

Special Parts

In addition to the standard electronic components, some special components are used:

- Components, manufactured or selected by Philips to meet specific performance requirements.
- Components which are important for the safety of the instrument, marked with 'S' in the parts list.

ATTENTION:

Both type of components may only be replaced by components obtained through your local Philips organisation.

4.5.2. Mechanical parts, miscellaneous, electrical parts not on units

Item	Fig.	Quantity	Order number	Description	
01	5	1	5322 447 94324	cover, grey	
01	5	1	5322 447 90395	cover, brown	
02	5	4	5322 462 44174	foot (bottom side), grey	
02	5	4	5322 462 10222	foot (bottom side), brown	
03	5	2	5322 520 34164	bearing bush	
04	5	2	5322 530 84075	spring	
05	5	2	5322 528 34101	rachet	
06	5	2	5322 532 54425	ring for handle, grey	
06	5	2	5322 532 51481	ring for handle, brown	
07	5	2	5322 498 54048	arm for handle	
08	5	1	5322 498 54051	carrying handle	
09	5	2	5322 414 64053	knob, grey	
09	5	2	5322 414 30043	knob, brown	
11	5	1	5322 447 94188	back frame	
12	5	4	5322 462 44176	foot (rear side)	
13	_	1	5322 502 14164	coin-slot screw (rear side)	
14	_	1	4822 530 70124	locking washer (rear side)	
15	5	1	5322 414 74042	cover for knob(scale), grey	
15	5	1	5322 414 70043	cover for knob(scale), brown	
16	5	1	5322 414 70043	cover for knob(switch), grey	
16	5	1	5322 414 74015	cover for knob(switch), brown	
17	5	3	5322 414 70013	cover for knob, grey	
17	5	3	5322 414 74014	cover for knob, brown	
18	5	3	5322 414 74015	cover for small knob, grey	
18	5	3	5322 414 70016	cover for small knob, brown	
19	5	1	5322 414 70016	front frame	
21	6	1	5322 401 14275	cable clamp	*s
22	6	1	5322 321 14048	mains cable 1850	*S
23	6	1	5322 321 14048	lead through	*S
24	6	1	4822 253 30012	fuse 200 mAT	*S
566	6	1	5322 121 44028		*S
26	6	3	5322 405 94178	27N 2x2N7 250 V, Line filter	"5
29	5	22		print holder	
29	5	22	5322 414 25851 5322 414 20033	knob for pushbutton, grey	
30	5	1	5322 414 20033	knob for pushbutton, brown	
30	5	1		knob for switch, grey	
31	5	3	5322 414 30071	knob for switch, brown	
31			5322 414 34083	knob, grey	
31	5 5	3	5322 414 30069	knob, brown	
32	5	2	5322 414 34075	knob small (4mm axle), grey	
33	5	2	5322 414 30053	knob small (4mm axle), brown	
33	J	1	5322 414 34096	knob small (1/8" axle), grey	
	E	1	5322 414 30041	knob small (1/8" axle), brown	
34	5	1	5322 414 44076	dial mounted, grey	
34	5	1	5322 414 40024	dial mounted, brown	
35	5	1	5322 414 74047	locating mark, grey	
35	5	1	5322 414 70044	locating mark, brown	
370-875	5	6	5322 267 10004	BNC-socket	
251	5/6	1	5322 276 14393	mains switch	* S

^{*}S = safety component

*****S

Item	Fig.	Quantity	Order number	Description
				
252/1/2/4	4	3	5322 276 10959	pushbutton switch
252/3/5	4	2	5322 276 10961	pushbutton switch
252/6	4	1	5322 276 14221	pushbutton switch
253/1-3	7	3	5322 276 14221	pushbutton switch
253/4/7-16	7	11	5322 276 14271	pushbutton switch
253/5/6	7	2	5322 276 10961	pushbutton switch
38	6	1	5322 273 64059	rotary switch
39	-	1	5322 405 74027	print holder unit 2
40	7	5	5322 255 44265	heat sink
41	7	2	5322 255 44037	heat sink
42	_	_	5322 390 24013	silicon paste DC 340
648	6	1	5322 105 24014	potmeter 5KO/5
673	6	1	5322 103 64043	potmeter 5KO/5
601,647	6	2	4822 101 20441	carbon potm. 10K lin.
646	6	1	4822 101 20416	carbon potm. 4K7 lin.
821	6	1	4822 101 20431	carbon potm. 2K2 lin.
822	6	1	5322 101 64029	carbon potm. 22K lin.+switch
751	6	1	5322 146 24486	mains transformer

*S = Safety component

4.5.3. Electrical parts

351

Some parts are listed in chapter 4.5.2.

INTEGRATED CIRCUITS				
404	5322	209	85512	IC MC1458N
401	5322		85254	
402			86056	
404,405	5322		85254	
406-408			85565	
410			86349	
4.11				
412	5322	209	84778	10 000011
TRANSISTORS				
301,302	4822	130	44197	BC558B
303	5322		44509	
304,305			40937	~
304,303			41095	
315, 316	4822		44 19 7	
317			40937	
318,323			41024	
310,319			44594	
311,312	4822		40937	
313			44197	
314,324			40937	
320,322			44196	
321			44594	
325,328			40937	
326,327	5322	130	44034	2N2219A
329			40937	
330			40021	
331-333	1822	130	44197	BC558B
334-336	4822	130	40937	BC548B
337,350	4822	130	44197	BC558B

4822 130 40937 BC548B

352,354	5322 130 40468	2N2905A
353,355	5322 130 44034	2N2219A
356	4822 130 40959	BC 547B
357	4822 130 44568	BC 557B
358	5322 130 44034	2N 2219A
359	5322 130 40468	2N 2905A
DIODES		
421	5322 130 34321	1N4151 (BAW62)
422,432	4822 130 34297	BZX79-C10
423,424	5322 130 34321	1N4151 (BAW62)
425,426	4822 130 34278	BZX79-C6V8
427,431	5322 130 34321	1N4151 (BAW62)
428,430 429 433,434 435,436 437,438	4822 130 34278 4822 130 30229 5322 130 34321 4822 130 34278 4822 130 34174	BZX79-B6V8 AAZ 15 1N4151 (BAW62) BZX79-C6V8 BZX79-C4V7
439-447	4822 130 34195	BZX79-C13
443	4822 130 30862	BZX79-B9C1
444	4822 130 34167	BZX79-B6V2
445,446	5322 130 34321	1N4151 (BAW62)
447	4822 130 34047	BZX-C1V5
451	5322 130 32031	RECTIFIER SKB2/08L5A
452,453	4822 130 34195	BZX79-B13
454	4822 130 34048	BZX75-C2V8
455,456	5322 130 34321	1N4151 (BAW62)
COILS		
471,472	5322 158 14018	33 MUH
473	5322 158 10276	4,7 MUH
474	5322 158 10283	150 MUH

CAPACITORS

ITEM	ORDERING NUMBER	FARAD	TOL (%)	VOLTS	REMARKS
501 502,521 503,522 504,523	5322 121 40323 5322 121 40197 5322 121 54213 5322 121 40283	100H 1 MU 6MU8 3MU3	5 5 5 5	100 . 100 . 100 .	POLYESTER FOIL POLYESTER FOIL POLYSTYRENE FOIL POLYESTER FOIL
505 506,529 508,507 509	4822 122 30027 5322 124 24202 4822 122 30128 4822 122 30114	1N 2MU2 4N7 2N2	2 10 10	100 . 63 . 100 . 100 .	CERAMIC PLATE ELECTRULYTIC CERAMIC PLATE CERAMIC PLATE
511,512 513,514 515 516	4822 124 20499 4822 122 31067 5322 121 40323 5322 125 54068	22MU 33P 100N 2-22P	2 10	40 . 100 . 100 .	ELECTROLYTIC CERAMIC PLATE POLYESTER FOIL TRIMMER
517 518 519 520	4822 122 31074 5322 121 54085 5322 121 54154 5322 121 54124 4822 122 31175	56P 976P 10N 100N 1N	2 1 1 1 10	100 . 125 . 63 . 100 .	CERAMIC PLATE POLYSTYRENE FOIL POLYSTYRENE FOIL POLYSTYRENE FOIL CERAMIC PLATE
524,525 526,528 527 530 531	4822 122 30103 4822 122 31074 5322 124 24202 4822 122 31178	22N 56P 2MU2 680P	-20+80 2	63 . 100 . 63 . 100 .	CERAMIC PLATE CERAMIC PLATE ELECTROLYTIC CERAMIC PLATE
532 533	5322 121 40324 4822 122 30128	15N 4N7	10 10	250 . 100 .	POLYESTER FOIL CERAMIC PLATE
540 542,551 543,544	5322 121 44138 5322 121 40323 4822 124 20798	47N 100N 3300MU	20 20	250 . 100 . 40 .	POLYESTER FOIL POLYESTER FOIL ELECTROLYTIC

ITEM 545,546 547,548 550,552 555,565 557,558 559,560 561,562 566	ORDERING NUMBER 4822 124 20583 4822 124 20499 4822 122 30103 4822 122 31054 4822 122 31069 4822 125 50062 4822 122 30128 4822 122 30103 5322 122 30108 4822 122 31063	OHM 1MU 22MU 22MU 10P 39P 2-10P 4N7 22N 100N 22P	TOL (%) -20+80 2 2 10 -20+80 10 2	TYPE 63 . 40 . 63 . 100 . 100 . 100 . 50 .	REMARKS ELECTROLYTIC ELECTROLYTIC CERAMIC PLATE CERAMIC PLATE CERAMIC PLATE TRIMMER CERAMIC PLATE CERAMIC PLATE CERAMIC PLATE POLYESTER FOIL CERAMIC PLATE
RESISTORS					
ITEM	ORDERING NUMBER	ОНМ	TOL (%)	TYPE	REMARKS
602,623 603,614 604 6006 6007 6008 6011 6612 6112 6115 616 617,618 621,629 621,623 624,636 625,626 627 633 633,642	4822 100 10029 5322 116 54619 5322 116 55458 5322 116 55276 4822 100 10037 5322 116 54661 5322 116 546629 5322 116 50672 5322 116 55258 5322 116 54686 5322 116 54686 5322 116 54643 5322 116 54649 5322 116 54649 5322 116 54649 5322 116 54643 5322 116 54643 5322 116 54679 5322 116 54643 5322 116 54643 5322 116 54643 5322 116 54643 5322 116 54643 5322 116 54643 5322 116 54643 5322 116 54643 5322 116 54643	2K2 106 K 1224K 134K 14K6 14K6 14K6 14K6 14K6 14K6 14K6 14	1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	LIN 55555	TRIMMING POTM METAL FILM
641	5322 116 50608	6K19	1	MR25	METAL FILM
644 645,666 649 650 651 652 653,660 * 6566 657 658 659,661 663,670 710 6667 668 667,719 676,677 676,677 676,677 678 681,682 683,684 687,723 686 687,688 687,688 687,688 687,688 687,688 687,688 687,688 687,688 687,688 688,688 687,688 687,688 688,688 687,688 687,688 687,688 687,688 687,688 688,688 687,688 687,688 688,688 687,688 688,688 688,688 687,688 688,688	5322 116 54619 5322 116 54619 5322 116 55304 5322 116 55508 5322 116 55508 5322 116 55038 5322 116 55038 5322 116 55038 5322 116 55038 5322 116 55038 5322 116 55038 5322 116 55038 5322 116 55328 4 5322 116 54619 4822 100 10035 4822 100 10035 4822 100 10035 4822 116 54557 4822 100 10038 5322 116 54557 4822 100 10038 5322 116 54557 4822 100 10038 5322 116 54557 5322 116 54585 5322 116 54585 5322 116 50588 5322 116 50588 5322 116 50588 5322 116 50588 5322 116 50588 5322 116 50588 5322 116 50588	6K81 10046K41 938K64 938K61 122K0 122K0 10K 147K61	1 1 0,25 0,25 0,25 1 1 1 1 1 1 1 1 1 1 1 1	MR20555555555555555555555555555555555555	METAL FILM TRIMMING POTM TRIMMING POTM TRIMMING POTM TRIMMING POTM TRIMMING POTM METAL FILM

ITEM	ORDERING NUMBER	ОНМ	TOL (%)	TYPE	REMARKS
ITEM 689,691 690,725 692,726 693,693 694,705 695,705 697,699 697,699 701 702 703 704 706 707 718,713 716 717 718,720 721 722 724 727 728 732,839 732,839 779	ORDERING NUMBER 5322 116 54606 5322 116 54567 5322 116 50581 5322 116 50581 5322 116 54539 5322 116 54539 5322 116 546414 5322 116 50669 5322 116 50767 5322 116 54446 5322 116 54446 5322 116 54646 4822 100 10051 4822 101 10108 5322 116 54529 5322 116 54529 5322 116 54529 5322 116 54514 5322 116 54514 5322 116 54514 5322 116 55451 5322 116 54696 5322 116 54576 5322 116 54576 5322 116 54576 5322 116 50524 5322 116 54576 5322 116 50524 5322 116 50524 5322 116 50524 5322 116 50524 5322 116 50524 5322 116 50524	OHM 1599 118058522147511205852214762200215852200215865220021586522002166533748032200216653374803241354200216654665220020000000000000000000000000000	TOL (%) 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	TYMERENE	REMARKS METAL FILM METAL FIL
780 781,784 782,785 783,813 807,8819 808,812 811 815,817 816 818 820 821 823 824 825 828,829 832,833 834,836 827,833 834,836 837,838 840,841 843,844 845,846 845,846 845,846 845,846 845,846 845,846 845,846 845,846 845,846 845,846 845,846	5322 116 50728 4822 100 10075 5322 116 54525 5322 116 50509 5322 116 54743 5322 116 54549 4822 100 10038 5322 116 50414 4822 100 10038 5322 116 50414 4822 100 10079 5322 116 50669 5322 116 50669 5322 116 50679 4822 100 10079 5322 116 554686 5322 116 554882 5322 116 54624 5322 116 54624 5322 116 54624 5322 116 54624 5322 116 54624 5322 116 55482 5322 116 54624 5322 116 54624 5322 116 54624 5322 116 54624 5322 116 54624 5322 116 54624 5322 116 54624 5322 116 54624 5322 116 54624 5322 116 54624 5322 116 54624 5322 116 54624 5322 116 54458 5322 116 54458 5322 116 54458 5322 116 54458 5322 116 54458 5322 116 54458	1 K 8 7 1 1 1 1 8 7 1 1 1 1 8 7 1 1 1 1 8 7 1 1 1 1	1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	MR25	METAL FILM TRIMMING POTM METAL FILM METAL FILM METAL FILM TRIMMING POTM METAL FILM TRIMMING POTM METAL FILM TRIMMING POTM METAL FILM TRIMMING POTM METAL FILM

4.3.4. Table of checks and adjustments

2.4. 1	
Section Sect	
Controlled current source, main oscillator Section	
20	
20	
1	
1	
1	
1	
1	
1	
1	en pos. and neg. value
20	
1	- feet this
1,	
20	15 <i>2</i> 0
20	10, 20
1 20	
2 4	2527 and repeat seq. 2.10.1, 2
20	
20	
20 x x x x x x x x x	
20	
2	
10	
20 x	
20 x	
20 X X	be adjustable within 5 4
10	. 01/
10 x x x x x x x x x	to U V
Veep control	
20 x	
20	
20 x x x x x x x x x x x x x x x x x x x	
20 x	
20	triangular wave
	<u> </u>
20 X	
20 A T 500 ms +50 ms	
20 X T 5s ±0,5s	
20 ^ single sweep	nust be performed
f running sweep	must be reset to tSTART
g. 20 y x x x x f pushing the bu	tton, frequency must keep fter H/L going edge


Seq	۱.	FREQUENCY WAVE	AMPLITUDE DC	DUTY CYCL	LE SWEEP	measuring point			
		dial OFFSET range FORM Comparison Compa	10 dB 30 dB 30 dB 1/2 max max 50Ω/600Ω 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0	% % % % % % % % % % % % % % % % % % %	10	(front) (rear) point	adjustment	meas. quantity	measured value remark

4. sine shaper, amplifier, attenuator

4.1.	10	×			x					x				x	•	-		x						×						VDC	-5 ±0,5 V +5 ± 0,5 V	
4.2.	5	х				>	(×					x		×			x						х				and changes	55	7	optimal waveform	rise time ≤ 75 n sec.
4.3.	10	X		x			x						x		x			x						×					80 81	8 K	< 0,3 %	
4.4.	10	X		x			x						x		x			х						х					81	6 ADC	0 ±20 mV	
4.5.	20	x				>	(x							x	x				x	х			х	×			С		81	4 V _{pp}		optimize frequency response
4.6.	5					>	x							x	x			х						х						K	<2 %	
4.7.	10	×		x			•	•••	-0 -0	•••				×	×			X						×		,				V _{pp}	15 V ±1 V 7,5 ±0,5 V	
4.8.	10	х		×			x				•	••		×	x			x						×						atten	nom. value ±0,2 dB	
4.9.	10	×		×			L _X L					>	<		x			x						х						V_{pp}	1,5 ±0,3 V	
4.10.	10	х		x			x)	(x	·x			х			П	-		х						Vpp	2,3 ±0,15 V	
4.11.	10	х		×			x						x		x			×								x				Vp	4,5 5 V 0 ± 0,2 V	TTL-output
4.12.	0.4	×		×			x						x		×			×							D	x				V _{DC}	0 +5 V 100 Hz 2 kHz	

C: Connect SWEEP OUT to external x-input of oscilloscope

D: DC voltage to be fed in










Fig. 2 Front view
Frontansicht
Face avant




Fig.4. Unit 2, component lay-out




Fig.6. Bottom view





Fig. 7. Unit 1, component lay-out






Fig.8. Overall circuit-diagram PM5132